

DEC 24 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130135
Name	Michael Potter	ID#	5811 (P.2001) (214)
		Date	12/20/08
A	Agency	Alaska State Troopers	Phone # 656-1634
	Instrument Location	Galena	
B	Alco S/N	77675	Target Value .076
			High Pressure N/A
	Alco Test Values	.074	.069
		1 st Alco	2 nd Alco
	Signature	[Signature]	
		(OVER)	

Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Michael Potter, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

1/6/09
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130135 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130135

DECEMBER 20, 2008 ✓

OPERATOR'S NAME: ✓

POTTER/MICHAEL/J ✓

OPERATOR'S NUMBER: 5041

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:

X

O.L. #: X

DEPT/AGENCY: GALO

CASE/REPORT: X

TEST TYPE: U

ALCO TARGET VALUE: .076 ✓

ALCO S/N: 77675

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 30.33 in

ALCO TARGET: .077 15:45

BLANK TEST .000 15:46

INTERNAL STANDARD VERIFIED 15:46

ALCO TV 30.33 in .074 ✓ 15:47

BLANK TEST .000 15:47

SUBJECT SAMPLE .000 15:48

BLANK TEST .000 15:49

ALCO TV 30.33 in .069 ✓ 15:49

BLANK TEST .000 15:50

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130135

DECEMBER 20, 2008 ✓

TIME: 15:37

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 30.33 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&*()-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxy z{|}~